

14B NCAC 16 .0603 P.S.E. INSTRUMENTS

An instrument used for P.S.E. examinations shall be capable of measuring and recording voice reactions on a graph. This recording must be in a form suitable for examination by another P.S.E. examiner. Such recordings shall be available to the Board or its designated representative. The instrument must be given maintenance, cleaning, adjustment, and demagnetizing periodically as recommended by the manufacturer and not less than once after each eight hours of continuous mechanical operation. The examiner shall conduct a test pattern of the instrument prior to each examination.

*History Note: Authority G.S. 74C-5;
Eff. June 1, 1984;
Transferred and Recodified from 12 NCAC 07D .0603 Eff. July 1, 2015;
Readopted Eff. August 1, 2020.*